

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070905.D
 Acq On : 2 Feb 2022 21:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 03 05:58:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	50.000	52.996	-6.0	95	0.00
3 P	Chloromethane	50.000	50.956	-1.9	97	0.00
4 C	Vinyl Chloride	50.000	38.244	23.5#	68	0.00
5 T	Bromomethane	50.000	50.555	-1.1	98	0.00
6 T	Chloroethane	50.000	56.117	-12.2	102	0.00
7 T	Trichlorofluoromethane	50.000	53.590	-7.2	98	0.00
8 T	Diethyl Ether	50.000	52.055	-4.1	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.318	-4.6	94	0.00
10 T	Methyl Iodide	50.000	51.376	-2.8	91	0.00
11 T	Tert butyl alcohol	250.000	277.373	-10.9	101	0.00
12 CM	1,1-Dichloroethene	50.000	52.371	-4.7#	94	0.00
13 T	Acrolein	250.000	249.194	0.3	95	0.00
14 T	Allyl chloride	50.000	53.434	-6.9	95	0.00
15 T	Acrylonitrile	250.000	278.825	-11.5	100	0.00
16 T	Acetone	250.000	233.155	6.7	85	0.00
17 T	Carbon Disulfide	50.000	49.815	0.4	92	0.00
18 T	Methyl Acetate	50.000	55.540	-11.1	100	0.00
19 T	Methyl tert-butyl Ether	50.000	53.201	-6.4	94	0.00
20 T	Methylene Chloride	50.000	51.210	-2.4	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.813	-3.6	93	0.00
22 T	Diisopropyl ether	50.000	53.192	-6.4	96	0.00
23 T	Vinyl Acetate	250.000	269.213	-7.7	95	0.00
24 P	1,1-Dichloroethane	50.000	51.914	-3.8	94	0.00
25 T	2-Butanone	250.000	267.690	-7.1	96	0.00
26 T	2,2-Dichloropropane	50.000	48.617	2.8	88	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.074	-2.1	93	0.00
28 T	Bromochloromethane	50.000	52.346	-4.7	94	0.00
29 T	Tetrahydrofuran	250.000	280.080	-12.0	101	0.00
30 C	Chloroform	50.000	50.797	-1.6#	93	0.00
31 T	Cyclohexane	50.000	50.325	-0.7	95	0.00
32 T	1,1,1-Trichloroethane	50.000	51.364	-2.7	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.837	0.3	91	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	91	0.00
35 S	Dibromofluoromethane	50.000	49.566	0.9	90	0.00
36 T	1,1-Dichloropropene	50.000	51.269	-2.5	94	0.00
37 T	Ethyl Acetate	50.000	55.593	-11.2	100	0.00
38 T	Carbon Tetrachloride	50.000	50.833	-1.7	91	0.00
39 T	Methylcyclohexane	50.000	52.318	-4.6	92	0.00
40 TM	Benzene	50.000	50.722	-1.4	94	0.00
41 T	Methacrylonitrile	50.000	55.531	-11.1	100	0.00
42 TM	1,2-Dichloroethane	50.000	50.831	-1.7	93	0.00
43 T	Isopropyl Acetate	50.000	49.577	0.8	96	0.00
44 TM	Trichloroethene	50.000	50.129	-0.3	93	0.00
45 C	1,2-Dichloropropane	50.000	52.190	-4.4#	95	0.00
46 T	Dibromomethane	50.000	51.825	-3.7	93	0.00
47 T	Bromodichloromethane	50.000	51.856	-3.7	93	0.00
48 T	Methyl methacrylate	50.000	56.385	-12.8	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070905.D
 Acq On : 2 Feb 2022 21:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 03 05:58:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1106.857	-10.7	100	0.00
50 S	Toluene-d8	50.000	50.206	-0.4	92	0.00
51 T	4-Methyl-2-Pentanone	250.000	277.660	-11.1	99	0.00
52 CM	Toluene	50.000	49.999	0.0#	93	0.00
53 T	t-1,3-Dichloropropene	50.000	51.767	-3.5	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.339	-4.7	92	0.00
55 T	1,1,2-Trichloroethane	50.000	51.293	-2.6	94	0.00
56 T	Ethyl methacrylate	50.000	53.883	-7.8	94	0.00
57 T	1,3-Dichloropropane	50.000	52.069	-4.1	93	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	237.701	4.9	97	0.00
59 T	2-Hexanone	250.000	278.014	-11.2	98	0.00
60 T	Dibromochloromethane	50.000	51.634	-3.3	92	0.00
61 T	1,2-Dibromoethane	50.000	51.459	-2.9	92	0.00
62 S	4-Bromofluorobenzene	50.000	50.925	-1.8	92	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	91	0.00
64 T	Tetrachloroethene	50.000	50.141	-0.3	91	0.00
65 PM	Chlorobenzene	50.000	50.010	-0.0	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.052	-2.1	92	0.00
67 C	Ethyl Benzene	50.000	52.048	-4.1#	93	0.00
68 T	m/p-Xylenes	100.000	103.522	-3.5	93	0.00
69 T	o-Xylene	50.000	51.189	-2.4	92	0.00
70 T	Styrene	50.000	54.223	-8.4	93	0.00
71 P	Bromoform	50.000	51.238	-2.5	89	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	87	0.00
73 T	Isopropylbenzene	50.000	52.145	-4.3	93	0.00
74 T	N-ethyl acetate	50.000	54.704	-9.4	95	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.731	-3.5	95	0.00
76 T	1,2,3-Trichloropropane	50.000	49.724	0.6	88	0.00
77 T	Bromobenzene	50.000	50.207	-0.4	91	0.00
78 T	n-propylbenzene	50.000	52.535	-5.1	92	0.00
79 T	2-Chlorotoluene	50.000	51.582	-3.2	93	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.017	-6.0	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	54.042	-8.1	92	0.00
82 T	4-Chlorotoluene	50.000	51.787	-3.6	92	0.00
83 T	tert-Butylbenzene	50.000	52.183	-4.4	92	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.644	-5.3	92	0.00
85 T	sec-Butylbenzene	50.000	53.101	-6.2	93	0.00
86 T	p-Isopropyltoluene	50.000	53.834	-7.7	93	0.00
87 T	1,3-Dichlorobenzene	50.000	50.938	-1.9	91	0.00
88 T	1,4-Dichlorobenzene	50.000	49.348	1.3	91	0.00
89 T	n-Butylbenzene	50.000	51.627	-3.3	92	0.00
90 T	Hexachloroethane	50.000	52.785	-5.6	92	0.00
91 T	1,2-Dichlorobenzene	50.000	50.582	-1.2	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.404	-6.8	95	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.660	2.7	89	0.00
94 T	Hexachlorobutadiene	50.000	49.174	1.7	89	0.00
95 T	Naphthalene	50.000	48.313	3.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
Data File : VN070905.D
Acq On : 2 Feb 2022 21:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 03 05:58:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.480	1.0	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6