



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070904.D
 Acq On : 2 Feb 2022 20:16
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Title : SW846 8260
 Last Update : Wed Feb 02 15:30:39 2022



AutoFind: Scans 4004, 4005, 4006; Background Corrected with Scan 3987

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	32435	PASS
75	95	30	60	52.0	83901	PASS
95	95	100	100	100.0	161387	PASS
96	95	5	9	6.5	10560	PASS
173	174	0.00	2	0.8	968	PASS
174	95	50	100	76.1	122832	PASS
175	174	5	9	7.4	9095	PASS
176	174	95	101	97.0	119115	PASS
177	176	5	9	7.0	8315	PASS