

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020224\
 Data File : VN080873.D
 Acq On : 02 Feb 2024 14:54
 Operator : JC\MD
 Sample : VN0202MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0202MBL01

Quant Time: Feb 02 23:50:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	78935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	203791	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	212649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88002	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	70760	51.818	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.640%
35) Dibromofluoromethane	8.171	113	53887	43.220	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.440%
50) Toluene-d8	10.565	98	227341	46.587	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.180%
62) 4-Bromofluorobenzene	12.847	95	95841	56.360	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020224\
Data File : VN080873.D
Acq On : 02 Feb 2024 14:54
Operator : JC\MD
Sample : VN0202MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0202MBL01

Quant Time: Feb 02 23:50:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

