

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN020320\
 Data File : VN059964.D
 Acq On : 3 Feb 2020 10:40
 Operator : JC/MD
 Sample : VN0203MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0203MBL01

Quant Time: Feb 03 11:41:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	197834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	303643	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	269836	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	108291	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	111694	50.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.68%	
35) Dibromofluoromethane	7.56	113	92316	50.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.96%	
50) Toluene-d8	10.08	98	358907	51.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.32%	
62) 4-Bromofluorobenzene	12.40	95	118775	46.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020320\
Data File : VN059964.D
Acq On : 3 Feb 2020 10:40
Operator : JC/MD
Sample : VN0203MBL01
Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0203MBL01

Quant Time: Feb 03 11:41:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

