

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020320\
 Data File : VN059975.D
 Acq On : 3 Feb 2020 15:25
 Operator : JC/MD
 Sample : PB126526ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126526ZHE#10

Quant Time: Feb 04 04:06:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	193968	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	298684	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	279193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	124385	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	110930	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
35) Dibromofluoromethane	7.57	113	90478	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
50) Toluene-d8	10.08	98	357193	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
62) 4-Bromofluorobenzene	12.40	95	126734	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
Target Compounds						
20) Methylene Chloride	4.53	84	5577	1.609	ug/l #	80
43) Isopropyl Acetate	8.14	43	54127	12.899	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020320\
Data File : VN059975.D
Acq On : 3 Feb 2020 15:25
Operator : JC/MD
Sample : PB126526ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126526ZHE#10

Quant Time: Feb 04 04:06:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

