

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
 Data File : VN070922.D
 Acq On : 3 Feb 2022 11:37
 Operator : JC/MD
 Sample : VN0203MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0203MBL01

Quant Time: Feb 04 04:41:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	993093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1593664	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1397567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	526303	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	636181	47.326	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.660%
35) Dibromofluoromethane	8.021	113	463338	48.533	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.060%
50) Toluene-d8	10.440	98	1918113	48.933	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.731	95	629669	44.581	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
Data File : VN070922.D
Acq On : 3 Feb 2022 11:37
Operator : JC/MD
Sample : VN0203MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0203MBL01

Quant Time: Feb 04 04:41:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

