

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
 Data File : VN070926.D
 Acq On : 3 Feb 2022 13:34
 Operator : JC/MD
 Sample : N1362-04
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RO-341

Quant Time: Feb 04 04:42:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	954599	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1516053	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1270508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	550208	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	605443	46.856	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.720%
35) Dibromofluoromethane	8.016	113	432120	47.580	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.160%
50) Toluene-d8	10.440	98	1834447	49.194	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.380%
62) 4-Bromofluorobenzene	12.731	95	604676	45.003	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.000%
Target Compounds						
					Qvalue	
29) Tetrahydrofuran	7.705	42	10406	1.436	ug/l #	68
59) 2-Hexanone	11.089	43	18484	1.258	ug/l	98
93) 1,2,4-Trichlorobenzene	15.263	180	12186	1.199	ug/l	97
96) 1,2,3-Trichlorobenzene	15.700	180	13616	3.058	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
Data File : VN070926.D
Acq On : 3 Feb 2022 13:34
Operator : JC/MD
Sample : N1362-04
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RO-341

Quant Time: Feb 04 04:42:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

