

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
 Data File : VN070940.D
 Acq On : 3 Feb 2022 19:33
 Operator : JC/MD
 Sample : N1342-11DL 100X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31528DL2

Quant Time: Feb 04 04:45:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	966965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1527035	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1361694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	556718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	606749	46.356	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.720%		
35) Dibromofluoromethane	8.021	113	442564	48.379	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.760%		
50) Toluene-d8	10.443	98	1849296	49.236	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.480%		
62) 4-Bromofluorobenzene	12.731	95	653460	48.284	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.560%		
Target Compounds					Qvalue	
8) Diethyl Ether	3.824	74	17858	2.436	ug/l	88
11) Tert butyl alcohol	5.388	59	61478	22.353	ug/l #	84
16) Acetone	4.296	43	119180	15.228	ug/l	97
18) Methyl Acetate	4.865	43	27353	1.073	ug/l #	87
19) Methyl tert-butyl Ether	5.624	73	70886	1.905	ug/l #	89
25) 2-Butanone	7.345	43	72714	6.437	ug/l	98
40) Benzene	8.461	78	1114019	23.860	ug/l	99
52) Toluene	10.505	92	942256	32.572	ug/l	96
67) Ethyl Benzene	11.846	91	74744	1.439	ug/l	97
68) m/p-Xylenes	11.950	106	148051	7.691	ug/l	95
69) o-Xylene	12.280	106	54230	2.815	ug/l	92
84) 1,2,4-Trimethylbenzene	13.366	105	51481	1.337	ug/l	96
95) Naphthalene	15.507	128	59648	5.443	ug/l	98
96) 1,2,3-Trichlorobenzene	15.700	180	12642	2.946	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
Data File : VN070940.D
Acq On : 3 Feb 2022 19:33
Operator : JC/MD
Sample : N1342-11DL 100X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31528DL2

Quant Time: Feb 04 04:45:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

