

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
 Data File : VN070941.D
 Acq On : 3 Feb 2022 19:58
 Operator : JC/MD
 Sample : N1342-01DL 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31523DL2

Quant Time: Feb 04 04:45:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1002650	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1583173	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1404452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	537866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	629940	46.415	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.840%		
35) Dibromofluoromethane	8.024	113	452517	47.713	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.420%		
50) Toluene-d8	10.443	98	1914395	49.162	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.320%		
62) 4-Bromofluorobenzene	12.733	95	644835	45.957	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.920%		
Target Compounds						
11) Tert butyl alcohol	5.390	59	16228	5.690	ug/l #	75
16) Acetone	4.293	43	21401	2.637	ug/l	99
40) Benzene	8.461	78	932263	19.259	ug/l	99
52) Toluene	10.507	92	287527	9.587	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
Data File : VN070941.D
Acq On : 3 Feb 2022 19:58
Operator : JC/MD
Sample : N1342-01DL 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31523DL2

Quant Time: Feb 04 04:45:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

