

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
 Data File : VN070943.D
 Acq On : 3 Feb 2022 20:49
 Operator : JC/MD
 Sample : N1342-15DL 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31535DL2

Quant Time: Feb 04 04:46:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1015173	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1609041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1428341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	554554	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	640319	46.598	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.200%
35) Dibromofluoromethane	8.021	113	468195	48.573	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.140%
50) Toluene-d8	10.443	98	1954036	49.373	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.731	95	663913	46.556	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.120%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.385	59	24966	8.646	ug/l #	93
16) Acetone	4.293	43	62566	7.615	ug/l	99
40) Benzene	8.461	78	1671202	33.970	ug/l	99
52) Toluene	10.505	92	1136918	37.298	ug/l	97
67) Ethyl Benzene	11.846	91	67739	1.243	ug/l	98
68) m/p-Xylenes	11.950	106	141572	7.011	ug/l	93
69) o-Xylene	12.280	106	45190	2.236	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020322\
Data File : VN070943.D
Acq On : 3 Feb 2022 20:49
Operator : JC/MD
Sample : N1342-15DL 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31535DL2

Quant Time: Feb 04 04:46:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

