

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN020323\  
 Data File : VN076526.D  
 Acq On : 03 Feb 2023 12:16  
 Operator : JC\MD  
 Sample : VN0203WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0203WBL01

Quant Time: Feb 03 22:14:20 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N011623W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 17 03:59:22 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 8.230  | 168   | 534324   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.106  | 114   | 880597   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 11.865 | 117   | 748378   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.794 | 152   | 296193   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 8.582  | 65    | 304407   | 50.881   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | = 101.760% |          |
| 35) Dibromofluoromethane    | 8.171  | 113   | 265124   | 48.130   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 96.260%  |          |
| 50) Toluene-d8              | 10.571 | 98    | 1005789  | 49.440   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | = 98.880%  |          |
| 62) 4-Bromofluorobenzene    | 12.853 | 95    | 330011   | 49.691   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 99.380%  |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN020323\  
Data File : VN076526.D  
Acq On : 03 Feb 2023 12:16  
Operator : JC\MD  
Sample : VN0203WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0203WBL01

Quant Time: Feb 03 22:14:20 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N011623W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jan 17 03:59:22 2023  
Response via : Initial Calibration

