

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020323\
 Data File : VN076532.D
 Acq On : 03 Feb 2023 14:59
 Operator : JC\MD
 Sample : 01415-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Feb 03 22:18:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	593757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	977698	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	837915	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	331607	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	341981	51.440	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.880%
35) Dibromofluoromethane	8.171	113	297365	48.622	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.240%
50) Toluene-d8	10.571	98	1122900	49.715	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	12.853	95	366559	49.713	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020323\
Data File : VN076532.D
Acq On : 03 Feb 2023 14:59
Operator : JC\MD
Sample : 01415-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Feb 03 22:18:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

