

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085620.D
 Acq On : 03 Feb 2025 14:22
 Operator : JC\MD
 Sample : VN0203WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0203WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/04/2025
 Supervised By : Mahesh Dadoda 02/04/2025

Quant Time: Feb 04 00:58:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	316818	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	277592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132351	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	137375	47.978	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.960%		
35) Dibromofluoromethane	8.165	113	112154	51.027	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.060%		
50) Toluene-d8	10.565	98	403865	51.716	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.440%		
62) 4-Bromofluorobenzene	12.847	95	130381	48.807	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.620%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	36725	15.291	ug/l	96
3) Chloromethane	2.359	50	37630	14.471	ug/l	98
4) Vinyl Chloride	2.512	62	49451	18.920	ug/l	93
5) Bromomethane	2.959	94	32239	20.420	ug/l	96
6) Chloroethane	3.124	64	33310	20.102	ug/l	100
7) Trichlorofluoromethane	3.500	101	68134	17.966	ug/l	98
8) Diethyl Ether	3.965	74	23505	17.941	ug/l	89
9) 1,1,2-Trichlorotrifluo...	4.371	101	41678	19.512	ug/l	98
10) Methyl Iodide	4.594	142	46834	19.146	ug/l	95
11) Tert butyl alcohol	5.512	59	33754	102.951	ug/l	98
12) 1,1-Dichloroethene	4.347	96	34518	18.132	ug/l	83
13) Acrolein	4.183	56	22257	49.729	ug/l	96
14) Allyl chloride	5.024	41	51626	16.712	ug/l	92
15) Acrylonitrile	5.724	53	103456	99.342	ug/l	99
16) Acetone	4.424	43	83180	89.909	ug/l	94
17) Carbon Disulfide	4.712	76	82042	13.997	ug/l	98
18) Methyl Acetate	5.018	43	59280	21.072	ug/l	91
19) Methyl tert-butyl Ether	5.794	73	123329	19.951	ug/l	99
20) Methylene Chloride	5.271	84	44642	19.492	ug/l	91
21) trans-1,2-Dichloroethene	5.789	96	37520	18.442	ug/l	94
22) Diisopropyl ether	6.665	45	123987	18.083	ug/l	98
23) Vinyl Acetate	6.600	43	413834	86.245	ug/l	95
24) 1,1-Dichloroethane	6.565	63	79023	18.901	ug/l	97
25) 2-Butanone	7.482	43	128769	94.582	ug/l	95
26) 2,2-Dichloropropane	7.488	77	67615	20.004	ug/l	96
27) cis-1,2-Dichloroethene	7.488	96	47023	19.623	ug/l	93
28) Bromochloromethane	7.812	49	36218	18.620	ug/l	89
29) Tetrahydrofuran	7.841	42	81866	94.844	ug/l	91
30) Chloroform	7.965	83	85403	19.765	ug/l	97
31) Cyclohexane	8.253	56	56062	15.428	ug/l #	82
32) 1,1,1-Trichloroethane	8.165	97	71876	18.963	ug/l	97
36) 1,1-Dichloropropene	8.371	75	51010	16.536	ug/l	99
37) Ethyl Acetate	7.559	43	51943	16.682	ug/l #	97
38) Carbon Tetrachloride	8.359	117	62929	17.819	ug/l	99
39) Methylcyclohexane	9.594	83	47049	16.231	ug/l	97
40) Benzene	8.600	78	174408	18.817	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085620.D
 Acq On : 03 Feb 2025 14:22
 Operator : JC\MD
 Sample : VN0203WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0203WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/04/2025
 Supervised By : Mahesh Dadoda 02/04/2025

Quant Time: Feb 04 00:58:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	26968	16.648	ug/l #	88
42) 1,2-Dichloroethane	8.665	62	61132	17.512	ug/l	97
43) Isopropyl Acetate	8.682	43	91136	18.267	ug/l	97
44) Trichloroethene	9.353	130	39267	18.200	ug/l	99
45) 1,2-Dichloropropane	9.618	63	44301	18.712	ug/l	98
46) Dibromomethane	9.706	93	32808	19.205	ug/l	99
47) Bromodichloromethane	9.882	83	67537	19.406	ug/l	99
48) Methyl methacrylate	9.676	41	36989	16.475	ug/l #	87
49) 1,4-Dioxane	9.688	88	16068	424.978	ug/l	91
51) 4-Methyl-2-Pentanone	10.441	43	276729	95.585	ug/l	95
52) Toluene	10.629	92	107257	19.972	ug/l	97
53) t-1,3-Dichloropropene	10.835	75	63248	19.230	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	65503	18.645	ug/l	94
55) 1,1,2-Trichloroethane	11.012	97	44024	20.714	ug/l	97
56) Ethyl methacrylate	10.871	69	58187	17.556	ug/l	91
57) 1,3-Dichloropropane	11.159	76	74646	20.199	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	106559	79.007	ug/l	98
59) 2-Hexanone	11.194	43	196651	96.535	ug/l	95
60) Dibromochloromethane	11.359	129	51933	20.240	ug/l	98
61) 1,2-Dibromoethane	11.465	107	42700	20.183	ug/l	99
64) Tetrachloroethene	11.100	164	36100	19.076	ug/l	95
65) Chlorobenzene	11.888	112	119360	19.628	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	45386	20.335	ug/l	98
67) Ethyl Benzene	11.959	91	189154	19.098	ug/l	100
68) m/p-Xylenes	12.070	106	148722	40.628	ug/l	95
69) o-Xylene	12.394	106	68283	19.519	ug/l	96
70) Styrene	12.412	104	117567	20.307	ug/l	96
71) Bromoform	12.576	173	34300	21.478	ug/l #	97
73) Isopropylbenzene	12.694	105	172454	19.306	ug/l	99
74) N-amyl acetate	12.494	43	67688	16.859	ug/l	92
75) 1,1,2,2-Tetrachloroethane	12.935	83	66716	21.144	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	59990m	22.301	ug/l	
77) Bromobenzene	12.976	156	45507	19.499	ug/l	95
78) n-propylbenzene	13.035	91	208681	19.734	ug/l	99
79) 2-Chlorotoluene	13.123	91	132631	19.380	ug/l	97
80) 1,3,5-Trimethylbenzene	13.170	105	147177	19.947	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	20135	20.251	ug/l	90
82) 4-Chlorotoluene	13.217	91	131384	19.280	ug/l	98
83) tert-Butylbenzene	13.435	119	119331	19.259	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	146712	19.950	ug/l	98
85) sec-Butylbenzene	13.612	105	172226	20.055	ug/l	99
86) p-Isopropyltoluene	13.729	119	137244	19.216	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	83838	19.507	ug/l	99
88) 1,4-Dichlorobenzene	13.806	146	83412	18.724	ug/l	100
89) n-Butylbenzene	14.053	91	112690	18.291	ug/l	100
90) Hexachloroethane	14.335	117	31482	19.257	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	81715	19.061	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	11390	19.768	ug/l	94
93) 1,2,4-Trichlorobenzene	15.388	180	35819	17.977	ug/l	99
94) Hexachlorobutadiene	15.500	225	18661	17.641	ug/l	99
95) Naphthalene	15.641	128	107097	18.013	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	36123	17.917	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
Data File : VN085620.D
Acq On : 03 Feb 2025 14:22
Operator : JC\MD
Sample : VN0203WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0203WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025

Quant Time: Feb 04 00:58:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

