

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085624.D
 Acq On : 03 Feb 2025 16:00
 Operator : JC\MD
 Sample : PB166424ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166424ZHE#03

Quant Time: Feb 04 00:59:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	341580	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	322934	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	115512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	143006	53.418	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.840%			
35) Dibromofluoromethane	8.165	113	120069	50.668	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.340%			
50) Toluene-d8	10.565	98	446904	53.079	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.160%			
62) 4-Bromofluorobenzene	12.847	95	123090	42.738	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 85.480%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	33749	39.016	ug/l #	85
20) Methylene Chloride	5.271	84	3696	1.726	ug/l #	74
43) Isopropyl Acetate	8.683	43	147932	27.501	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
Data File : VN085624.D
Acq On : 03 Feb 2025 16:00
Operator : JC\MD
Sample : PB166424ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166424ZHE#03

Quant Time: Feb 04 00:59:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

