

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085629.D
 Acq On : 03 Feb 2025 18:01
 Operator : JC\MD
 Sample : PB166424ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166424ZHE#08

Quant Time: Feb 04 00:59:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	350437	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	322910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	112899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	144583	54.260	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.520%		
35) Dibromofluoromethane	8.165	113	120003	49.360	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.720%		
50) Toluene-d8	10.565	98	451815	52.306	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 104.620%		
62) 4-Bromofluorobenzene	12.847	95	124312	42.071	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 84.140%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	24942	28.970	ug/l	93
20) Methylene Chloride	5.271	84	2542	1.193	ug/l #	65
43) Isopropyl Acetate	8.688	43	184853	33.497	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
Data File : VN085629.D
Acq On : 03 Feb 2025 18:01
Operator : JC\MD
Sample : PB166424ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166424ZHE#08

Quant Time: Feb 04 00:59:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

