

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085632.D
 Acq On : 03 Feb 2025 19:13
 Operator : JC\MD
 Sample : Q1216-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-26.1-012825

Quant Time: Feb 04 01:00:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	339935	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288192	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	94832	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	137122	52.880	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.760%			
35) Dibromofluoromethane	8.165	113	107897	45.752	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.500%			
50) Toluene-d8	10.565	98	388253	46.336	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.680%			
62) 4-Bromofluorobenzene	12.847	95	108452	37.837	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 75.680%#			
Target Compounds						
16) Acetone	4.430	43	70347	83.961	ug/l	97
20) Methylene Chloride	5.283	84	3310	1.596	ug/l #	71
30) Chloroform	7.965	83	4833	1.235	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
Data File : VN085632.D
Acq On : 03 Feb 2025 19:13
Operator : JC\MD
Sample : Q1216-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-26.1-012825

Quant Time: Feb 04 01:00:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

