

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085634.D
 Acq On : 03 Feb 2025 20:02
 Operator : JC\MD
 Sample : Q1241-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-3.5-013025

Quant Time: Feb 04 02:58:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/04/2025
 Supervised By : Mahesh Dadoda 02/04/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.871	168	190736m	50.000	ug/l	-0.35
34) 1,4-Difluorobenzene	8.859	114	364470	50.000	ug/l	-0.24
63) Chlorobenzene-d5	11.812	117	339889m	50.000	ug/l	-0.05
72) 1,4-Dichlorobenzene-d4	13.776	152	119474	50.000	ug/l	-0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.259	65	167760	54.487	ug/l	-0.31
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.980%
35) Dibromofluoromethane	7.782	113	134150	53.055	ug/l	-0.38
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.100%
50) Toluene-d8	10.447	98	449937	50.083	ug/l	-0.12
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.160%
62) 4-Bromofluorobenzene	12.823	95	130014m	42.307	ug/l	-0.02
Spiked Amount	50.000	Range	77 - 121	Recovery	=	84.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
Data File : VN085634.D
Acq On : 03 Feb 2025 20:02
Operator : JC\MD
Sample : Q1241-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-3.5-013025

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025

Quant Time: Feb 04 02:58:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

