

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085637.D
 Acq On : 03 Feb 2025 21:13
 Operator : JC\MD
 Sample : Q1241-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-5.2-013025

Quant Time: Feb 04 01:01:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	330403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	315638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	139635	54.864	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.720%	
35) Dibromofluoromethane	8.165	113	113349	49.450	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.900%	
50) Toluene-d8	10.565	98	435589	53.485	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.980%	
62) 4-Bromofluorobenzene	12.847	95	130787	46.946	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 93.900%	
Target Compounds						Qvalue
16) Acetone	4.430	43	29252	35.571	ug/l	99
20) Methylene Chloride	5.277	84	5204	2.556	ug/l #	75
43) Isopropyl Acetate	8.688	43	130809	25.141	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
Data File : VN085637.D
Acq On : 03 Feb 2025 21:13
Operator : JC\MD
Sample : Q1241-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-5.2-013025

Quant Time: Feb 04 01:01:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

