

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085638.D
 Acq On : 03 Feb 2025 21:37
 Operator : JC\MD
 Sample : Q1241-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-5.4-013025

Quant Time: Feb 04 01:01:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	163991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	343832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	330622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	116237	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	144365	54.536	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.080%		
35) Dibromofluoromethane	8.165	113	119058	49.912	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.820%		
50) Toluene-d8	10.565	98	455608	53.758	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.520%		
62) 4-Bromofluorobenzene	12.847	95	132098	45.565	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 91.120%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	33701	39.402	ug/l	98
20) Methylene Chloride	5.277	84	4200	1.984	ug/l #	85
43) Isopropyl Acetate	8.688	43	130603	24.121	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
Data File : VN085638.D
Acq On : 03 Feb 2025 21:37
Operator : JC\MD
Sample : Q1241-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-5.4-013025

Quant Time: Feb 04 01:01:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

