

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
 Data File : VN085639.D
 Acq On : 03 Feb 2025 22:01
 Operator : JC\MD
 Sample : Q1241-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-51.4-013025

Quant Time: Feb 04 01:01:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	349686	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	332324	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	147102	54.109	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.220%		
35) Dibromofluoromethane	8.165	113	121025	49.887	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.780%		
50) Toluene-d8	10.565	98	452380	52.484	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 104.960%		
62) 4-Bromofluorobenzene	12.847	95	132884	45.069	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 90.140%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	24974	28.431	ug/l	96
20) Methylene Chloride	5.277	84	4678	2.151	ug/l	93
43) Isopropyl Acetate	8.688	43	113800	20.666	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020325\
Data File : VN085639.D
Acq On : 03 Feb 2025 22:01
Operator : JC\MD
Sample : Q1241-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-51.4-013025

Quant Time: Feb 04 01:01:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

