

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020419\
 Data File : VN053601.D
 Acq On : 4 Feb 2019 17:17
 Operator : JC/SP
 Sample : K1299-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-LEACH-AMND-COMP-2

Manual Integrations
 APPROVED

MMDadoda
 2/5/2019 9:54:19 AM

Quant Time: Feb 05 03:16:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	853506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1351667	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1199098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	497988	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	430859	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
35) Dibromofluoromethane	7.59	113	382255	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
50) Toluene-d8	10.09	98	1622966	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
62) 4-Bromofluorobenzene	12.40	95	529466	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
Target Compounds						
11) Tert butyl alcohol	4.79	59	6604m	13.032	ug/l	
16) Acetone	3.82	43	30310	13.801	ug/l	96
25) 2-Butanone	6.84	43	6684m	2.134	ug/l	
95) Naphthalene	15.13	128	59859	4.378	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020419\
Data File : VN053601.D
Acq On : 4 Feb 2019 17:17
Operator : JC/SP
Sample : K1299-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SPLP-LEACH-AMND-COMP-2

Manual Integrations
APPROVED

MMDadoda
2/5/2019 9:54:19 AM

Quant Time: Feb 05 03:16:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

