

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020419\
 Data File : VN053604.D
 Acq On : 4 Feb 2019 18:37
 Operator : JC/SP
 Sample : PB116782TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116782TB

Quant Time: Feb 05 03:21:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	845267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1362728	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1237517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	541176	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	438595	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
35) Dibromofluoromethane	7.59	113	431071	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
50) Toluene-d8	10.09	98	1639375	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
62) 4-Bromofluorobenzene	12.40	95	551237	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
Target Compounds						
16) Acetone	3.82	43	22561	10.373	ug/l	99
20) Methylene Chloride	4.56	84	54112	5.930	ug/l	98
25) 2-Butanone	6.85	43	5926	1.910	ug/l	93
43) Isopropyl Acetate	8.17	43	34216	2.582	ug/l #	88
95) Naphthalene	15.14	128	7546	2.043	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020419\
Data File : VN053604.D
Acq On : 4 Feb 2019 18:37
Operator : JC/SP
Sample : PB116782TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116782TB

Quant Time: Feb 05 03:21:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

