

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020419\
 Data File : VN053606.D
 Acq On : 4 Feb 2019 19:31
 Operator : JC/SP
 Sample : K1359-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1

Quant Time: Feb 05 03:36:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	863760	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1371479	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1233257	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	558210	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	435047	50.21	ug/l	0.00
Spiked Amount			Recovery	=	100.42%	
35) Dibromofluoromethane	7.59	113	421863	48.68	ug/l	0.00
Spiked Amount			Recovery	=	97.36%	
50) Toluene-d8	10.09	98	1640668	50.42	ug/l	0.00
Spiked Amount			Recovery	=	100.84%	
62) 4-Bromofluorobenzene	12.40	95	571561	52.67	ug/l	0.00
Spiked Amount			Recovery	=	105.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.78	59	102088	199.069	ug/l	98
16) Acetone	3.82	43	276216	124.274	ug/l	98
18) Methyl Acetate	4.32	43	112936	17.157	ug/l	99
20) Methylene Chloride	4.55	84	12154	1.303	ug/l	91
25) 2-Butanone	6.84	43	191904	60.538	ug/l	97
40) Benzene	8.04	78	507066	13.300	ug/l	98
52) Toluene	10.16	92	159994	7.071	ug/l	99
67) Ethyl Benzene	11.51	91	51369	1.194	ug/l	98
68) m/p-Xylenes	11.62	106	76938	4.685	ug/l	98
69) o-Xylene	11.95	106	33325	2.100	ug/l	98
70) Styrene	11.96	104	53526	2.081	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	82128	2.477	ug/l	88
95) Naphthalene	15.13	128	512459	21.909	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020419\
Data File : VN053606.D
Acq On : 4 Feb 2019 19:31
Operator : JC/SP
Sample : K1359-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
COMP-1

Quant Time: Feb 05 03:36:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

