

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020419\
 Data File : VN053607.D
 Acq On : 4 Feb 2019 19:57
 Operator : JC/SP
 Sample : K1359-03DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-3DL

Quant Time: Feb 05 03:39:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	841623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1335660	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1169453	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	487719	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	424494	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
35) Dibromofluoromethane	7.59	113	421027	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
50) Toluene-d8	10.09	98	1590311	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
62) 4-Bromofluorobenzene	12.40	95	514294	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
Target Compounds						
16) Acetone	3.82	43	6811	3.145	ug/l	90
18) Methyl Acetate	4.33	43	5534	0.863	ug/l #	89
25) 2-Butanone	6.84	43	80718	26.133	ug/l	99
52) Toluene	10.16	92	1066044	48.378	ug/l	100
68) m/p-Xylenes	11.62	106	8513	0.547	ug/l	94
95) Naphthalene	15.13	128	102280	6.344	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020419\
Data File : VN053607.D
Acq On : 4 Feb 2019 19:57
Operator : JC/SP
Sample : K1359-03DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
COMP-3DL

Quant Time: Feb 05 03:39:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

