

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
 Data File : VN065710.D
 Acq On : 4 Feb 2021 12:00
 Operator : JC/MD
 Sample : VN0204WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0204WBL01

Quant Time: Feb 05 02:14:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	191520	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	327142	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	284521	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	103083	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	115295	45.76	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.52%
35) Dibromofluoromethane	7.547	113	94865	48.93	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.86%
50) Toluene-d8	10.058	98	409673	53.76	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.52%
62) 4-Bromofluorobenzene	12.373	95	130787	47.66	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.32%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
Data File : VN065710.D
Acq On : 4 Feb 2021 12:00
Operator : JC/MD
Sample : VN0204WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0204WBL01

Quant Time: Feb 05 02:14:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

