

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
 Data File : VN065713.D
 Acq On : 4 Feb 2021 13:26
 Operator : JC/MD
 Sample : M1288-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-3S(ONSITE)

Quant Time: Feb 05 02:16:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	189896	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	321058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	277796	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	97030	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	114570	45.86	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.72%		
35) Dibromofluoromethane	7.550	113	92558	48.65	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.30%		
50) Toluene-d8	10.058	98	397995	53.22	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.44%		
62) 4-Bromofluorobenzene	12.373	95	124299	46.16	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	92.32%		
Target Compounds						
64) Tetrachloroethene	10.598	164	1860	0.80	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
Data File : VN065713.D
Acq On : 4 Feb 2021 13:26
Operator : JC/MD
Sample : M1288-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-3S(ONSITE)

Quant Time: Feb 05 02:16:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

