

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
 Data File : VN065721.D
 Acq On : 4 Feb 2021 16:43
 Operator : JC/MD
 Sample : M1288-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-4-PS

Quant Time: Feb 05 02:18:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	204563	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	345871	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	304869	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	103048	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	122381	45.47	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.94%
35) Dibromofluoromethane	7.550	113	99484	48.54	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.08%
50) Toluene-d8	10.058	98	428624	53.20	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.40%
62) 4-Bromofluorobenzene	12.373	95	137756	47.48	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.96%
Target Compounds						
30) Chloroform	7.332	83	1183	0.26	ug/l #	81
64) Tetrachloroethene	10.598	164	25805	10.10	ug/l	97
90) Hexachloroethane	13.843	117	5609	4.20	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
Data File : VN065721.D
Acq On : 4 Feb 2021 16:43
Operator : JC/MD
Sample : M1288-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PS

Quant Time: Feb 05 02:18:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

