

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
 Data File : VN065724.D
 Acq On : 4 Feb 2021 17:57
 Operator : JC/MD
 Sample : PB134422TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134422TB

Quant Time: Feb 05 02:19:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	191943	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	323647	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	288986	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	104262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	115665	45.80	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.60%
35) Dibromofluoromethane	7.547	113	93662	48.84	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.68%
50) Toluene-d8	10.058	98	404955	53.72	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.44%
62) 4-Bromofluorobenzene	12.373	95	133924	49.33	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.66%
Target Compounds						
16) Acetone	3.779	43	5349	5.44	ug/l	93
20) Methylene Chloride	4.496	84	7293	2.68	ug/l	98
43) Isopropyl Acetate	8.129	43	17506	3.16	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
Data File : VN065724.D
Acq On : 4 Feb 2021 17:57
Operator : JC/MD
Sample : PB134422TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134422TB

Quant Time: Feb 05 02:19:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

