

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
 Data File : VN065722.D
 Acq On : 4 Feb 2021 17:07
 Operator : JC/MD
 Sample : M1288-21
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-4-PD

Quant Time: Feb 05 02:19:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.624	168	200619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	343482	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	304495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	106035	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	121313	45.96	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.92%
35) Dibromofluoromethane	7.547	113	98098	48.20	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.40%
50) Toluene-d8	10.058	98	431295	53.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.82%
62) 4-Bromofluorobenzene	12.373	95	136542	47.39	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.78%
Target Compounds						
						Qvalue
44) Trichloroethene	8.798	130	959	0.37	ug/l	77
64) Tetrachloroethene	10.601	164	12095	4.74	ug/l	94
90) Hexachloroethane	13.846	117	797	0.58	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
Data File : VN065722.D
Acq On : 4 Feb 2021 17:07
Operator : JC/MD
Sample : M1288-21
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PD

Quant Time: Feb 05 02:19:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07




Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PD


Tgt Ion:168 Resp: 200619
Ion Ratio Lower Upper
168 100
99 58.0 42.6 63.8



#33
1,2-Dichloroethane-d4
Concen: 45.96 ug/l
RT: 7.987 min Scan# 1943
Delta R.T. 0.000 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07


Tgt Ion: 65 Resp: 121313
Ion Ratio Lower Upper
65 100
67 56.6 0.0 107.2








#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.550 min Scan# 2118
Delta R.T. 0.000 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PD







Tgt Ion:114 Resp: 343482
Ion Ratio Lower Upper
114 100
63 21.1 0.0 38.6
88 16.3 0.0 30.6



#35
Dibromofluoromethane
Concen: 48.20 ug/l
RT: 7.547 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07

Tgt Ion:113 Resp: 98098
Ion Ratio Lower Upper
113 100
111 102.3 82.2 123.2
192 18.2 15.4 23.0









#44
Trichloroethene
Concen: 0.37 ug/l
RT: 8.798 min Scan# 2195
Delta R.T. -0.000 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07




Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PD


Tgt Ion:130 Resp: 959
Ion Ratio Lower Upper
130 100
95 123.2 0.0 199.6



#50
Toluene-d8
Concen: 53.91 ug/l
RT: 10.058 min Scan# 2587
Delta R.T. 0.000 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07

Tgt Ion: 98 Resp: 431295
Ion Ratio Lower Upper
98 100
100 64.1 50.5 75.7









#62
4-Bromofluorobenzene
Concen: 47.39 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. 0.000 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07




Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PD


Tgt Ion: 95 Resp: 136542
Ion Ratio Lower Upper
95 100
174 70.8 0.0 169.0
176 67.6 0.0 162.4



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.376 min Scan# 2997
Delta R.T. 0.000 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07


Tgt Ion: 117 Resp: 304495
Ion Ratio Lower Upper
117 100
82 58.6 43.7 65.5
119 32.3 26.1 39.1








#64
Tetrachloroethene
Concen: 4.74 ug/l
RT: 10.601 min Scan# 2756
Delta R.T. 0.003 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07


Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PD





Tgt Ion:164 Resp: 12095
Ion Ratio Lower Upper
164 100
166 131.2 102.6 154.0
129 101.7 76.2 114.4
131 100.7 73.4 110.2




#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.315 min Scan# 3600
Delta R.T. 0.003 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07




Tgt Ion:152 Resp: 106035
Ion Ratio Lower Upper
152 100
115 59.3 28.5 85.6
150 157.1 0.0 347.6





#90
Hexachloroethane
Concen: 0.58 ug/l
RT: 13.846 min Scan# 3765
Delta R.T. 0.004 min
Lab File: VN065722.D
Acq: 4 Feb 2021 17:07

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PD



Tgt Ion:117 Resp: 797
Ion Ratio Lower Upper
117 100
201 72.1 37.8 113.3

