



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020421\
 Data File : VN065707.D
 Acq On : 4 Feb 2021 9:25
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Title : SW846 8260
 Last Update : Mon Jan 25 13:47:28 2021



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3294

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	2328	PASS
75	95	30	60	48.4	6252	PASS
95	95	100	100	100.0	12922	PASS
96	95	5	9	7.1	915	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.6	9250	PASS
175	174	5	9	7.6	707	PASS
176	174	95	101	96.9	8964	PASS
177	176	5	9	7.2	645	PASS