

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020425\
 Data File : VN085647.D
 Acq On : 04 Feb 2025 12:31
 Operator : JC\MD
 Sample : VN0204WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0204WBL01

Quant Time: Feb 05 01:02:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	197983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	379434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	328690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132287	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159588	49.936	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.880%
35) Dibromofluoromethane	8.165	113	132313	50.264	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.520%
50) Toluene-d8	10.565	98	460162	49.201	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.400%
62) 4-Bromofluorobenzene	12.847	95	141527	44.237	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020425\
Data File : VN085647.D
Acq On : 04 Feb 2025 12:31
Operator : JC\MD
Sample : VN0204WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0204WBL01

Quant Time: Feb 05 01:02:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

