

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020425\
 Data File : VN085658.D
 Acq On : 04 Feb 2025 17:50
 Operator : JC\MD
 Sample : Q1268-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-REF-5

Quant Time: Feb 05 01:09:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	213738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	402251	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	341688	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	134799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	151577	43.933	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.860%
35) Dibromofluoromethane	8.165	113	132536	47.493	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.980%
50) Toluene-d8	10.565	98	474290	47.835	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.680%
62) 4-Bromofluorobenzene	12.847	95	154147	45.448	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020425\
Data File : VN085658.D
Acq On : 04 Feb 2025 17:50
Operator : JC\MD
Sample : Q1268-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-5

Quant Time: Feb 05 01:09:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

