

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020425\
 Data File : VN085659.D
 Acq On : 04 Feb 2025 18:14
 Operator : JC\MD
 Sample : Q1268-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Feb 05 01:09:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	194531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	369397	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	314164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	129208	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	143797	45.793	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.580%
35) Dibromofluoromethane	8.165	113	122181	47.677	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.360%
50) Toluene-d8	10.565	98	435716	47.853	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.700%
62) 4-Bromofluorobenzene	12.847	95	140430	45.087	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020425\
Data File : VN085659.D
Acq On : 04 Feb 2025 18:14
Operator : JC\MD
Sample : Q1268-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Feb 05 01:09:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

