

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020519\
 Data File : VN053613.D
 Acq On : 5 Feb 2019 12:15
 Operator : JC/SP
 Sample : K1346-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-020419-A

Quant Time: Feb 06 04:05:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	824492	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1315469	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1187374	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	492523	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	423141	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
35) Dibromofluoromethane	7.59	113	410476	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
50) Toluene-d8	10.09	98	1574507	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	12.40	95	520166	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
Target Compounds						
16) Acetone	3.82	43	29235	13.780	ug/l	99
20) Methylene Chloride	4.55	84	5619359	631.321	ug/l	99
25) 2-Butanone	6.85	43	5328	1.761	ug/l	90
43) Isopropyl Acetate	8.17	43	29727	2.324	ug/l #	85
95) Naphthalene	15.13	128	22850	2.756	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020519\
Data File : VN053613.D
Acq On : 5 Feb 2019 12:15
Operator : JC/SP
Sample : K1346-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
CL-02-020419-A

Quant Time: Feb 06 04:05:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

