

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020521\
 Data File : VN065738.D
 Acq On : 5 Feb 2021 15:17
 Operator : JC/MD
 Sample : M1314-04DL 500X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 40966DL

Quant Time: Feb 05 22:48:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	188899	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	319878	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	285715	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	108351	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	111972	45.06	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.12%
35) Dibromofluoromethane	7.547	113	93134	49.13	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.26%
50) Toluene-d8	10.058	98	399118	53.56	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.12%
62) 4-Bromofluorobenzene	12.373	95	129925	48.42	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.84%

Target Compounds						Qvalue
16) Acetone	3.779	43	298172	308.20	ug/l	97
52) Toluene	10.122	92	61009	10.48	ug/l	99
67) Ethyl Benzene	11.482	91	10501	0.94	ug/l	99
68) m/p-Xylenes	11.588	106	17024	4.19	ug/l	96
69) o-Xylene	11.916	106	7966	2.02	ug/l	99
78) n-propylbenzene	12.563	91	4263	0.38	ug/l	95
80) 1,3,5-Trimethylbenzene	12.704	105	5742	0.71	ug/l	96
84) 1,2,4-Trimethylbenzene	13.013	105	20756	2.58	ug/l	98
95) Naphthalene	15.100	128	7827	4.49	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020521\
Data File : VN065738.D
Acq On : 5 Feb 2021 15:17
Operator : JC/MD
Sample : M1314-04DL 500X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
40966DL

Quant Time: Feb 05 22:48:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

