

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020521\
 Data File : VN065741.D
 Acq On : 5 Feb 2021 16:32
 Operator : JC/MD
 Sample : M1318-03 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31408-31410

Quant Time: Feb 05 22:49:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	190973	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	322447	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	285193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	104330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	113342	45.11	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.22%
35) Dibromofluoromethane	7.547	113	93448	48.91	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.82%
50) Toluene-d8	10.058	98	402978	53.65	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.30%
62) 4-Bromofluorobenzene	12.373	95	131219	48.52	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.04%
Target Compounds						
18) Methyl Acetate	4.290	43	4155	1.52	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020521\
Data File : VN065741.D
Acq On : 5 Feb 2021 16:32
Operator : JC/MD
Sample : M1318-03 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31408-31410

Quant Time: Feb 05 22:49:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

