

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020521\
 Data File : VN065742.D
 Acq On : 5 Feb 2021 16:56
 Operator : JC/MD
 Sample : M1318-08 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31394-5-31403-31387

Quant Time: Feb 05 22:49:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	196916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	333698	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	294365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	104573	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	115039	44.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.82%
35) Dibromofluoromethane	7.550	113	95944	48.52	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.04%
50) Toluene-d8	10.058	98	414552	53.33	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.66%
62) 4-Bromofluorobenzene	12.373	95	133153	47.57	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.14%
Target Compounds						
52) Toluene	10.122	92	6817	1.12	ug/l	98
84) 1,2,4-Trimethylbenzene	13.010	105	14992	1.93	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020521\
Data File : VN065742.D
Acq On : 5 Feb 2021 16:56
Operator : JC/MD
Sample : M1318-08 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31394-5-31403-31387

Quant Time: Feb 05 22:49:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

