


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020521\
Data File : VN065728.D
Acq On : 5 Feb 2021 10:39
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Title : SW846 8260
Last Update : Mon Jan 25 13:47:28 2021



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	4399	PASS
75	95	30	60	45.6	12430	PASS
95	95	100	100	100.0	27277	PASS
96	95	5	9	6.5	1774	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	68.2	18603	PASS
175	174	5	9	7.9	1477	PASS
176	174	95	101	95.8	17824	PASS
177	176	5	9	6.6	1171	PASS