

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
 Data File : VN080897.D
 Acq On : 05 Feb 2024 15:26
 Operator : JC\MD
 Sample : P1371-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T60-AOP

Quant Time: Feb 06 04:43:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	80177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	192741	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	203951	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	66903	48.234	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.460%
35) Dibromofluoromethane	8.165	113	50941	43.200	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.400%
50) Toluene-d8	10.571	98	205486	44.523	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.040%
62) 4-Bromofluorobenzene	12.847	95	92343	57.416	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.840%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	7828	15.561	ug/l	95
30) Chloroform	7.959	83	3622	1.566	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
Data File : VN080897.D
Acq On : 05 Feb 2024 15:26
Operator : JC\MD
Sample : P1371-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T60-AOP

Quant Time: Feb 06 04:43:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

