

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
 Data File : VN080900.D
 Acq On : 05 Feb 2024 16:38
 Operator : JC\MD
 Sample : P1371-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T240-AOP

Quant Time: Feb 06 04:44:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	80648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	195501	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	212597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90094	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	67452	48.346	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.700%
35) Dibromofluoromethane	8.165	113	52648	44.017	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.040%
50) Toluene-d8	10.571	98	225982	48.272	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.540%
62) 4-Bromofluorobenzene	12.847	95	89723	55.000	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.000%
Target Compounds						
16) Acetone	4.442	43	4852	9.589	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
Data File : VN080900.D
Acq On : 05 Feb 2024 16:38
Operator : JC\MD
Sample : P1371-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T240-AOP

Quant Time: Feb 06 04:44:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

