

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
 Data File : VN080907.D
 Acq On : 05 Feb 2024 19:26
 Operator : JC\MD
 Sample : P1366-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-1-0-2

Quant Time: Feb 06 04:47:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	78342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	189672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	212569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	68634	50.641	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.280%	
35) Dibromofluoromethane	8.165	113	49587	42.732	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 85.460%	
50) Toluene-d8	10.571	98	233444	51.399	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.800%	
62) 4-Bromofluorobenzene	12.853	95	89360	56.461	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.920%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	14914	30.342	ug/l #	77
20) Methylene Chloride	5.283	84	6663	5.164	ug/l	92
25) 2-Butanone	7.488	43	5423	5.809	ug/l	92
37) Ethyl Acetate	7.553	43	4885	2.200	ug/l #	94
43) Isopropyl Acetate	8.688	43	122142	30.215	ug/l #	89
52) Toluene	10.629	92	3861	1.062	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
Data File : VN080907.D
Acq On : 05 Feb 2024 19:26
Operator : JC\MD
Sample : P1366-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-1-0-2

Quant Time: Feb 06 04:47:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

