

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
 Data File : VN080909.D
 Acq On : 05 Feb 2024 20:14
 Operator : JC\MD
 Sample : P1367-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-6-0-2

Quant Time: Feb 06 04:48:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	75382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	190094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	221477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	78346	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	68563	52.575	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.160%			
35) Dibromofluoromethane	8.171	113	52022	44.731	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 89.460%			
50) Toluene-d8	10.571	98	223150	49.023	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.040%			
62) 4-Bromofluorobenzene	12.853	95	91685	57.801	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 115.600%			
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	21102	44.617	ug/l	99
18) Methyl Acetate	5.024	43	3390	2.057	ug/l #	78
20) Methylene Chloride	5.277	84	6540	5.268	ug/l #	82
25) 2-Butanone	7.494	43	7739	8.615	ug/l	94
37) Ethyl Acetate	7.547	43	3863	1.736	ug/l #	85
43) Isopropyl Acetate	8.688	43	121069	29.883	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
Data File : VN080909.D
Acq On : 05 Feb 2024 20:14
Operator : JC\MD
Sample : P1367-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-6-0-2

Quant Time: Feb 06 04:48:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

