

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
 Data File : VN080910.D
 Acq On : 05 Feb 2024 20:38
 Operator : JC\MD
 Sample : P1367-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-7-0-2

Quant Time: Feb 06 04:48:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	77222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	186239	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	204801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	66180	49.539	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.080%
35) Dibromofluoromethane	8.171	113	51118	44.863	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.720%
50) Toluene-d8	10.571	98	208422	46.736	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.480%
62) 4-Bromofluorobenzene	12.853	95	84169	54.161	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.320%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	132402	273.274	ug/l	96
20) Methylene Chloride	5.271	84	7782	6.119	ug/l #	67
25) 2-Butanone	7.489	43	9173	9.968	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
Data File : VN080910.D
Acq On : 05 Feb 2024 20:38
Operator : JC\MD
Sample : P1367-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-7-0-2

Quant Time: Feb 06 04:48:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

