

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
 Data File : VN080911.D
 Acq On : 05 Feb 2024 21:02
 Operator : JC\MD
 Sample : P1367-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-6-7-9

Quant Time: Feb 06 04:49:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	80388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	191299	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	205824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	86246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	84938	61.076	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.160%
35) Dibromofluoromethane	8.171	113	56396	48.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	10.571	98	214499	46.826	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.660%
62) 4-Bromofluorobenzene	12.847	95	86498	54.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.380%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	14123	28.001	ug/l	97
20) Methylene Chloride	5.277	84	5392	4.073	ug/l #	79
25) 2-Butanone	7.500	43	6152	6.422	ug/l	96
43) Isopropyl Acetate	8.694	43	128242	31.454	ug/l #	87
52) Toluene	10.630	92	38269	10.439	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
Data File : VN080911.D
Acq On : 05 Feb 2024 21:02
Operator : JC\MD
Sample : P1367-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-6-7-9

Quant Time: Feb 06 04:49:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

