

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
 Data File : VN080913.D
 Acq On : 05 Feb 2024 21:50
 Operator : JC\MD
 Sample : P1367-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-7-5-7D

Quant Time: Feb 06 04:50:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 02/06/2024
 Supervised By : Mahesh Dadoda 02/06/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	79484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	173067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	200035	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	82550	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	64791	47.119	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 94.240%			
35) Dibromofluoromethane	8.171	113	56051	52.937	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.880%			
50) Toluene-d8	10.571	98	211158	50.953	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.900%			
62) 4-Bromofluorobenzene	12.853	95	86944	60.205	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 120.400%			
Target Compounds						
						Qvalue
16) Acetone	4.436	43	14294	28.663	ug/l	92
20) Methylene Chloride	5.265	84	6769m	5.171	ug/l	
25) 2-Butanone	7.489	43	5396	5.697	ug/l	91
37) Ethyl Acetate	7.571	43	4085	2.016	ug/l #	84
43) Isopropyl Acetate	8.694	43	115589	31.338	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020524\
Data File : VN080913.D
Acq On : 05 Feb 2024 21:50
Operator : JC\MD
Sample : P1367-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-7-5-7D

Quant Time: Feb 06 04:50:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/06/2024
Supervised By :Mahesh Dadoda 02/06/2024

