

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN020525\  
 Data File : VN085688.D  
 Acq On : 05 Feb 2025 19:35  
 Operator : JC\MD  
 Sample : Q1289-03  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :


Quant Time: Feb 06 00:47:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N011425W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jan 15 02:16:08 2025  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|--------|----------|
| -----                         |        |       |          |          |        |          |
| Internal Standards            |        |       |          |          |        |          |
| 1) Pentafluorobenzene         | 8.224  | 168   | 206383   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene       | 9.100  | 114   | 404979   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5          | 11.865 | 117   | 373277   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4    | 13.788 | 152   | 142618   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds   |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4     | 8.577  | 65    | 160819   | 48.273   | ug/l   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 74 - 125 | Recovery | =      | 96.540%  |
| 35) Dibromofluoromethane      | 8.165  | 113   | 135100   | 48.086   | ug/l   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 75 - 124 | Recovery | =      | 96.180%  |
| 50) Toluene-d8                | 10.565 | 98    | 516501   | 51.742   | ug/l   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 86 - 113 | Recovery | =      | 103.480% |
| 62) 4-Bromofluorobenzene      | 12.847 | 95    | 154152   | 45.144   | ug/l   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 77 - 121 | Recovery | =      | 90.280%  |
| Target Compounds              |        |       |          |          |        |          |
|                               |        |       |          |          | Qvalue |          |
| 11) Tert butyl alcohol        | 5.518  | 59    | 192      | 0.503    | ug/l # | 71       |
| 16) Acetone                   | 4.430  | 43    | 36860    | 34.243   | ug/l   | 96       |
| 18) Methyl Acetate            | 5.024  | 43    | 2889     | 0.883    | ug/l   | 99       |
| 20) Methylene Chloride        | 5.265  | 84    | 3130     | 1.175    | ug/l   | 92       |
| 25) 2-Butanone                | 7.494  | 43    | 3500     | 2.210    | ug/l # | 59       |
| 30) Chloroform                | 7.953  | 83    | 1164     | 0.232    | ug/l # | 58       |
| 31) Cyclohexane               | 8.224  | 56    | 5330     | 1.261    | ug/l # | 5        |
| 36) 1,1-Dichloropropene       | 8.224  | 75    | 18565    | 4.708    | ug/l # | 51       |
| 37) Ethyl Acetate             | 7.559  | 43    | 3530     | 0.887    | ug/l # | 77       |
| 38) Carbon Tetrachloride      | 8.218  | 117   | 22799    | 5.050    | ug/l # | 18       |
| 42) 1,2-Dichloroethane        | 8.682  | 62    | 964      | 0.216    | ug/l # | 76       |
| 43) Isopropyl Acetate         | 8.688  | 43    | 164831   | 25.846   | ug/l # | 86       |
| 45) 1,2-Dichloropropane       | 9.735  | 63    | 1933     | 0.639    | ug/l # | 43       |
| 48) Methyl methacrylate       | 9.741  | 41    | 127753   | 44.513   | ug/l # | 45       |
| 49) 1,4-Dioxane               | 9.735  | 88    | 110      | 2.276    | ug/l # | 1        |
| 51) 4-Methyl-2-Pentanone      | 10.441 | 43    | 5433     | 1.468    | ug/l   | 92       |
| 53) t-1,3-Dichloropropene     | 10.724 | 75    | 966      | 0.230    | ug/l # | 89       |
| 54) cis-1,3-Dichloropropene   | 10.353 | 75    | 46130    | 10.272   | ug/l # | 64       |
| 56) Ethyl methacrylate        | 10.976 | 69    | 1789     | 2.307    | ug/l # | 1        |
| 57) 1,3-Dichloropropane       | 11.200 | 76    | 6569     | 1.391    | ug/l # | 43       |
| 59) 2-Hexanone                | 11.200 | 43    | 84958    | 32.627   | ug/l # | 57       |
| 71) Bromoform                 | 12.847 | 173   | 1207     | 0.562    | ug/l # | 1        |
| 74) N-amyl acetate            | 12.476 | 43    | 1276     | 0.295    | ug/l # | 42       |
| 76) 1,2,3-Trichloropropane    | 12.847 | 75    | 81052    | 27.961   | ug/l # | 1        |
| 81) trans-1,4-Dichloro-2-b... | 12.847 | 75    | 81052    | 75.651   | ug/l # | 3        |
| 95) Naphthalene               | 15.641 | 128   | 4282     | 0.668    | ug/l # | 95       |
| -----                         |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.224 min Scan# 1066  
Delta R.T. 0.006 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35




Instrument :  
MSVOA\_N  
ClientSampleId :


Tgt Ion:168 Resp: 206383  
Ion Ratio Lower Upper  
168 100  
99 63.5 53.6 80.4



#11  
Tert butyl alcohol  
Concen: 0.503 ug/l  
RT: 5.518 min Scan# 606  
Delta R.T. 0.006 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Tgt Ion: 59 Resp: 192  
Ion Ratio Lower Upper  
59 100  
57 0.0 8.6 13.0#









#16  
Acetone  
Concen: 34.243 ug/l  
RT: 4.430 min Scan# 41  
Delta R.T. 0.012 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35




Instrument :  
MSVOA\_N  
ClientSampleId :


Tgt Ion: 43 Resp: 36860  
Ion Ratio Lower Upper  
43 100  
58 32.0 23.8 35.6



#18  
Methyl Acetate  
Concen: 0.883 ug/l  
RT: 5.024 min Scan# 522  
Delta R.T. 0.006 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35


Tgt Ion: 43 Resp: 2889  
Ion Ratio Lower Upper  
43 100  
74 23.0 18.6 28.0








#20  
Methylene Chloride  
Concen: 1.175 ug/l  
RT: 5.265 min Scan# 51  
Delta R.T. -0.006 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Instrument :  
MSVOA\_N  
ClientSampleId :







| Tgt Ion | 84    | Resp: | 3130  |
|---------|-------|-------|-------|
| Ion     | Ratio | Lower | Upper |
| 84      | 100   |       |       |
| 49      | 120.8 | 104.1 | 156.1 |
| 51      | 38.2  | 31.0  | 46.4  |
| 86      | 56.5  | 51.9  | 77.9  |



#25  
2-Butanone  
Concen: 2.210 ug/l  
RT: 7.494 min Scan# 942  
Delta R.T. 0.018 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

| Tgt Ion | 43    | Resp: | 3500  |
|---------|-------|-------|-------|
| Ion     | Ratio | Lower | Upper |
| 43      | 100   |       |       |
| 72      | 46.2  | 20.2  | 30.4# |





#30

Chloroform

Concen: 0.232 ug/l

RT: 7.953 min Scan# 1066

Delta R.T. -0.006 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35

Instrument :

MSVOA\_N





ClientSampleId :

Tgt Ion: 83 Resp: 1164

Ion Ratio Lower Upper

83 100

85 31.7 51.8 77.6#



#31

Cyclohexane

Concen: 1.261 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.029 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35




Tgt Ion: 56 Resp: 5330


Ion Ratio Lower Upper

56 100

69 152.0 22.2 33.4#

84 125.6 66.4 99.6#





#33

1,2-Dichloroethane-d4

Concen: 48.273 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.006 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35

Instrument :

MSVOA\_N





ClientSampleId :

Tgt Ion: 65 Resp: 160819

Ion Ratio Lower Upper

65 100

67 52.8 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35




Tgt Ion: 114 Resp: 404979


Ion Ratio Lower Upper

114 100

63 20.1 0.0 47.6


88 15.1 0.0 32.6








#35  
Dibromofluoromethane  
Concen: 48.086 ug/l  
RT: 8.165 min Scan# 1056  
Delta R.T. 0.006 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Instrument :  
MSVOA\_N  
ClientSampleId :






Tgt Ion:113 Resp: 135100  
Ion Ratio Lower Upper  
113 100  
111 102.2 82.7 124.1  
192 17.7 14.3 21.5




#36  
1,1-Dichloropropene  
Concen: 4.708 ug/l  
RT: 8.224 min Scan# 1066  
Delta R.T. -0.141 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Tgt Ion: 75 Resp: 18565  
Ion Ratio Lower Upper  
75 100  
110 9.3 16.5 49.5#  
77 0.0 24.4 36.6#







#37

Ethyl Acetate

Concen: 0.887 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.006 min


Lab File: VN085688.D

Acq: 05 Feb 2025 19:35




Instrument :

MSVOA\_N

ClientSampleId :



|             |       |             |
|-------------|-------|-------------|
| Tgt Ion: 43 | Resp: | 3530        |
| Ion         | Ratio | Lower Upper |
| 43          | 100   |             |
| 61          | 4.0   | 10.9 16.3#  |
| 70          | 1.5   | 7.5 11.3#   |



#38

Carbon Tetrachloride

Concen: 5.050 ug/l




RT: 8.218 min Scan# 1065


Delta R.T. -0.141 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35

|              |       |             |
|--------------|-------|-------------|
| Tgt Ion: 117 | Resp: | 22799       |
| Ion          | Ratio | Lower Upper |
| 117          | 100   |             |
| 119          | 6.9   | 75.4 113.2# |
| 121          | 0.0   | 24.6 37.0#  |





#42

1,2-Dichloroethane

Concen: 0.216 ug/l

RT: 8.682 min Scan# 1144

Delta R.T. 0.018 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35

Instrument :

MSVOA\_N





ClientSampleId :

Tgt Ion: 62 Resp: 964

Ion Ratio Lower Upper

62 100

98 0.0 0.0 17.0



#43

Isopropyl Acetate

Concen: 25.846 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.006 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35




Tgt Ion: 43 Resp: 164831


Ion Ratio Lower Upper

43 100

61 16.3 20.7 31.1#

87 11.3 9.8 14.8





#45

1,2-Dichloropropane

Concen: 0.639 ug/l

RT: 9.735 min Scan# 11

Delta R.T. 0.118 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35

Instrument :

MSVOA\_N





ClientSampleId :

Tgt Ion: 63 Resp: 1933

Ion Ratio Lower Upper

63 100

65 0.0 25.6 38.4#



#48

Methyl methacrylate

Concen: 44.513 ug/l

RT: 9.741 min Scan# 1324

Delta R.T. 0.064 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35




Tgt Ion: 41 Resp: 127753


Ion Ratio Lower Upper

41 100

69 0.0 64.7 97.1#

39 56.0 49.0 73.6





#49

1,4-Dioxane

Concen: 2.276 ug/l

RT: 9.735 min Scan# 11

Delta R.T. 0.047 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35

Instrument :

MSVOA\_N

ClientSampleId :





Tgt Ion: 88 Resp: 110

Ion Ratio Lower Upper

88 100

43 1152013.6 26.6 39.8#

58 6293.6 59.5 89.3#



#50

Toluene-d8

Concen: 51.742 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085688.D




Acq: 05 Feb 2025 19:35


Tgt Ion: 98 Resp: 516501

Ion Ratio Lower Upper

98 100


100 63.9 52.2 78.4








#51  
4-Methyl-2-Pentanone  
Concen: 1.468 ug/l  
RT: 10.441 min Scan# 1443  
Delta R.T. 0.000 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Instrument :  
MSVOA\_N  
ClientSampleId :







Tgt Ion: 43 Resp: 5433  
Ion Ratio Lower Upper  
43 100  
58 42.8 30.5 45.7



#53  
t-1,3-Dichloropropene  
Concen: 0.230 ug/l  
RT: 10.724 min Scan# 1491  
Delta R.T. -0.112 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Tgt Ion: 75 Resp: 966  
Ion Ratio Lower Upper  
75 100  
77 35.3 23.5 35.3#









#54  
cis-1,3-Dichloropropene  
Concen: 10.272 ug/l  
RT: 10.353 min Scan# 1428  
Delta R.T. 0.047 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35




Instrument :  
MSVOA\_N  
ClientSampleId :


Tgt Ion: 75 Resp: 46130  
Ion Ratio Lower Upper  
75 100  
77 0.6 25.0 37.4#  
39 36.4 43.1 64.7#



#56  
Ethyl methacrylate  
Concen: 2.307 ug/l  
RT: 10.976 min Scan# 1534  
Delta R.T. 0.105 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Tgt Ion: 69 Resp: 1789  
Ion Ratio Lower Upper  
69 100  
41 3686.4 54.6 82.0#  
39 1754.1 32.4 48.6#





#57

1,3-Dichloropropane

Concen: 1.391 ug/l

RT: 11.200 min Scan# 1572

Delta R.T. 0.041 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35

Instrument :

MSVOA\_N





ClientSampleId :

Tgt Ion: 76 Resp: 6569

Ion Ratio Lower Upper

76 100

78 0.0 25.6 38.4#



#59

2-Hexanone

Concen: 32.627 ug/l

RT: 11.200 min Scan# 1572

Delta R.T. 0.012 min

Lab File: VN085688.D




Acq: 05 Feb 2025 19:35


Tgt Ion: 43 Resp: 84958

Ion Ratio Lower Upper

43 100

58 22.3 26.2 78.6#









#62  
4-Bromofluorobenzene  
Concen: 45.144 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35




Instrument :  
MSVOA\_N  
ClientSampleId :

Tgt Ion: 95 Resp: 154152  
Ion Ratio Lower Upper  
95 100  
174 75.8 0.0 145.0  
176 71.9 0.0 142.4




#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1685  
Delta R.T. -0.000 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Tgt Ion: 117 Resp: 373277  
Ion Ratio Lower Upper  
117 100  
82 55.2 48.6 72.8  
119 31.7 26.6 39.8







#71

Bromoform

Concen: 0.562 ug/l

RT: 12.847 min Scan# 1806

Delta R.T. 0.270 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35

Instrument :

MSVOA\_N

ClientSampleId :





Tgt Ion:173 Resp: 1207

Ion Ratio Lower Upper

173 100

175 580.3 24.4 73.2#

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN085688.D

Acq: 05 Feb 2025 19:35




Tgt Ion:152 Resp: 142618


Ion Ratio Lower Upper

152 100

115 60.7 31.1 93.3


150 159.0 0.0 343.6








#74  
N-amy1 acetate  
Concen: 0.295 ug/l  
RT: 12.476 min Scan# 11  
Delta R.T. -0.012 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Instrument :  
MSVOA\_N  
ClientSampleId :







|             |             |
|-------------|-------------|
| Tgt Ion: 43 | Resp: 1276  |
| Ion Ratio   | Lower Upper |
| 43 100      |             |
| 70 0.0      | 34.0 51.0#  |
| 55 0.0      | 21.4 32.2#  |
| 61 0.0      | 19.0 28.4#  |



#76  
1,2,3-Trichloropropane  
Concen: 27.961 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.141 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

|             |              |
|-------------|--------------|
| Tgt Ion: 75 | Resp: 81052  |
| Ion Ratio   | Lower Upper  |
| 75 100      |              |
| 77 1.3      | 109.7 329.2# |









#81  
trans-1,4-Dichloro-2-butene  
Concen: 75.651 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. 0.112 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Instrument :  
MSVOA\_N  
ClientSampleId :

Tgt Ion: 75 Resp: 81052  
Ion Ratio Lower Upper  
75 100  
53 0.1 95.6 143.4#  
89 0.1 37.0 55.6#



#95  
Naphthalene  
Concen: 0.668 ug/l  
RT: 15.641 min Scan# 2327  
Delta R.T. 0.006 min  
Lab File: VN085688.D  
Acq: 05 Feb 2025 19:35

Tgt Ion: 128 Resp: 4282  
Ion Ratio Lower Upper  
128 100  
127 14.2 10.6 16.0  
129 7.9 8.8 13.2#

