

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
 Data File : VN085673.D
 Acq On : 05 Feb 2025 13:35
 Operator : JC\MD
 Sample : PB166529ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166529ZHE#02

Quant Time: Feb 06 00:40:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	207997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	392588	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	374010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	141681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159821	47.601	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.200%
35) Dibromofluoromethane	8.165	113	133777	49.118	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.240%
50) Toluene-d8	10.565	98	513250	53.039	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.080%
62) 4-Bromofluorobenzene	12.847	95	152860	46.178	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.360%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	35950	33.139	ug/l	98
20) Methylene Chloride	5.277	84	3772	1.405	ug/l	91
43) Isopropyl Acetate	8.688	43	143908	23.277	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
Data File : VN085673.D
Acq On : 05 Feb 2025 13:35
Operator : JC\MD
Sample : PB166529ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166529ZHE#02

Quant Time: Feb 06 00:40:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

