

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
 Data File : VN085674.D
 Acq On : 05 Feb 2025 13:59
 Operator : JC\MD
 Sample : PB166529ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166529ZHE#03

Quant Time: Feb 06 00:40:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	197142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	385821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	355607	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132983	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	153871	48.352	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.700%
35) Dibromofluoromethane	8.165	113	128858	48.142	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	10.565	98	490580	51.585	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.180%
62) 4-Bromofluorobenzene	12.847	95	142193	43.709	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.420%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	33177	32.266	ug/l	93
20) Methylene Chloride	5.277	84	3488	1.370	ug/l #	77
43) Isopropyl Acetate	8.688	43	150023	24.692	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
Data File : VN085674.D
Acq On : 05 Feb 2025 13:59
Operator : JC\MD
Sample : PB166529ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166529ZHE#03

Quant Time: Feb 06 00:40:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

