

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
 Data File : VN085676.D
 Acq On : 05 Feb 2025 14:47
 Operator : JC\MD
 Sample : PB166529ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166529ZHE#05

Quant Time: Feb 06 00:41:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	190051	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	371968	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	352444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130271	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150115	48.932	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.860%
35) Dibromofluoromethane	8.159	113	124406	48.209	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.420%
50) Toluene-d8	10.565	98	483336	52.716	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.440%
62) 4-Bromofluorobenzene	12.847	95	142897	45.561	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.120%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	34361	34.665	ug/l	99
20) Methylene Chloride	5.277	84	3240	1.320	ug/l #	84
43) Isopropyl Acetate	8.688	43	164935	28.157	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
Data File : VN085676.D
Acq On : 05 Feb 2025 14:47
Operator : JC\MD
Sample : PB166529ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166529ZHE#05

Quant Time: Feb 06 00:41:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

