

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
 Data File : VN085678.D
 Acq On : 05 Feb 2025 15:35
 Operator : JC\MD
 Sample : PB166529ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166529ZHE#07

Quant Time: Feb 06 00:42:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	187268	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	375643	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	349588	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	151701	50.184	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.360%		
35) Dibromofluoromethane	8.159	113	127145	48.789	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.580%		
50) Toluene-d8	10.565	98	486253	52.516	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 105.040%		
62) 4-Bromofluorobenzene	12.847	95	142347	44.942	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 89.880%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	34592	35.416	ug/l	92
20) Methylene Chloride	5.277	84	3394	1.404	ug/l #	86
43) Isopropyl Acetate	8.688	43	175876	29.732	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
Data File : VN085678.D
Acq On : 05 Feb 2025 15:35
Operator : JC\MD
Sample : PB166529ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166529ZHE#07

Quant Time: Feb 06 00:42:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

